

CREATINE KINASE (CK-MB) ENZYME IMMUNOASSAY TEST KIT

OxisResearch Catalog Number: 11210

Enzyme Immunoassay for the Quantitative Determination of Creatine Kinase (CK-MB) in Human Serum

FOR RESEARCH USE ONLY

Store at 2 to 8°C.

PROPRIETARY AND COMMON NAMES

CK-MB Enzyme Immunoassay

INTENDED USE

For the quantitative determination of the CK-MB concentration in human serum.

INTRODUCTION

Creatine Kinase (CK-MB) is the enzyme being used as the definitive serum marker for the diagnosis or exclusion of acute myocardial infarction (AMI)¹⁻³.

The determination of CK-MB mass has proven to be more specific for myocardial necrosis than the long-standing CK-MB activity and CK-MB inhibition assays⁴⁻⁷.

CK-MB, released after AMI, is detectable in blood as early as 3-4 hours after the onset of symptoms, and remains elevated for approximately 65 hours post infarct⁸⁻⁹. CK-MB mass levels are reportedly 50% diagnostic for AMI after 3 hours and > 90% diagnostic at 6 hours¹⁰. Such accuracy makes CK-MB mass determinations useful in confirming AMI in patients presenting to the ER with non-diagnostic ECGs > 6 hours after the onset of symptoms¹¹⁻¹².

PRINCIPLE OF THE TEST

The CK-MB ELISA test is based on the principle of a solid phase enzyme-linked immunosorbent assay¹³⁻¹⁴. The assay system utilizes a monoclonal antibody directed against a distinct antigenic determinant on the CK-MB molecule is used for solid phase immobilization (on the microtiter wells). A goat anti-CK-MM antibody conjugated to horseradish peroxidase (HRP) is in the antibody-enzyme conjugate solution. The test sample is allowed to react simultaneously with the two antibodies, resulting in the CK-MB molecules being sandwiched between the solid phase and enzyme-linked antibodies. After a 1 hour incubation at room temperature, the wells are washed with water to remove unbound labeled antibodies. A solution of TMB Reagent is added and incubated at room temperature for 20 minutes, resulting in the development of a blue color. The color development is stopped with the addition of Stop Solution changing the color to yellow. The concentration of CK-MB is directly proportional to the color intensity of the test sample. Absorbance is measured spectrophotometrically at 450 nm.

REAGENTS

Materials provided with the kit:

- Antibody-coated microtiter plate with 96 wells.
- Liquid CK-MB standards containing; 0, 7.5, 15, 50, 100, and 200 ng/ml CK-MB. 1.0 ml for each standard dose. Store at -20°C or below.
- Enzyme Conjugate Reagent, 22 ml.
- TMB Reagent (One-Step), 11 ml.
- Stop Solution (1N HCl), 11 ml.

Materials required but not provided:

- Precision pipettes: 0.02 ml, 0.1 ml, 0.2 ml, and 1 ml.
- Disposable pipette tips.
- Distilled water.
- Vortex mixer or equivalent.
- Absorbent paper or paper towel.
- Graph paper.
- Microtiter plate reader.

SPECIMEN COLLECTION AND PREPARATION

Serum should be prepared from a whole blood specimen obtained by acceptable medical techniques. This kit is for use with serum samples without additives only.

STORAGE OF TEST KIT AND INSTRUMENTATION

1. CK-MB standards are not stable at room temperature. The standards are stable at 2-8°C for at least 7 days. It is recommended that the standards are aliquoted and stored at -20°C or below.
2. Unopened test kits should be stored at 2-8°C upon receipt and the microtiter plate should be kept in a sealed bag with desiccants to minimize exposure to damp air. Opened test kits will remain stable until the expiration date shown, provided it is stored as described above. A microtiter plate reader with a bandwidth of 10 nm or less and an optical density range of 0-2 OD or greater at 450 nm wavelength is acceptable for use in absorbance measurement.

REAGENT PREPARATION

1. All reagents should be brought to room temperature (18-25°C) before use.

ASSAY PROCEDURE

1. Secure the desired number of coated wells in the holder.
2. Dispense 20 µl of standard, specimens, and controls into appropriate wells.
3. Dispense 200 µl of Enzyme Conjugate Reagent to each well.
4. Thoroughly mix for 30 seconds. It is very important to have a complete mixing in this setup.
5. Incubate at room temperature (18-25°C) for 60 minutes.
6. Remove the incubation mixture by emptying plate content into a waste container.

7. Rinse and empty the microtiter wells 5 times with distilled or deionized water. (Please do not use tap water.)
8. Strike the wells sharply onto absorbent paper or paper towels to remove all residual water droplets.
9. Dispense 100 μ l of TMB Reagent into each well. Gently mix for 5 seconds.
10. Incubate at room temperature for 20 minutes.
11. Stop the reaction by adding 100 μ l of Stop Solution to each well.
12. Gently mix for 30 seconds. ***It is important to make sure that all the blue color changes to yellow color completely.***
13. Read the optical density at 450 nm with a microtiter plate reader within 15 minutes.


CALCULATION OF RESULTS

1. Calculate the average absorbance values (A_{450}) for each set of reference standards, control, and samples.
2. Construct a standard curve by plotting the mean absorbance obtained for each reference standard against its concentration in ng/ml on linear graph paper, with absorbance on the vertical (y) axis and concentration on the horizontal (x) axis.
3. Using the mean absorbance value for each sample, determine the corresponding concentration of CK-MB in ng/ml from the standard curve.
4. Any values obtained for diluted samples must be further converted by applying the appropriate dilution factor in the calculation.

EXAMPLE OF STANDARD CURVE

Results of a typical standard run with optical density readings at 450 nm shown in the Y axis against CK-MB concentrations shown in the X axis. This standard curve is for the purpose of illustration only, and should not be used to calculate unknowns. Each user should obtain his or her own data and standard curve.

CK-MB (ng/ml)	Absorbance (450 nm)
0	0.067
7.5	0.211
15	0.335
50	0.909
100	1.621
200	2.794



EXPECTED VALUES AND SENSITIVITY

Normal range for CK-MB reported by various literatures is between 0-9.0 ng/ml. It is recommended that each laboratory establish its own normal range. The minimum detectable concentration of CK-MB by this assay is estimated to be 2.5 ng/ml.

LIMITATIONS OF THE PROCEDURE

1. Reliable and reproducible results will be obtained when the assay procedure is carried out with a complete understanding of the package insert instructions and with adherence to good laboratory practice.
2. The wash procedure is critical. Insufficient washing will result in poor precision and falsely elevated absorbance readings.
3. Serum samples demonstrating gross lipemia, gross hemolysis, or turbidity should not be used with this test.
4. The results obtained from the use of this kit should be used only as an adjunct to other diagnostic procedures and information available to the physician.

REFERENCES

- 1 Lee, T.H., Goldman, L., Serum enzyme assays in the diagnosis of acute myocardial infarction. *Ann. Intern. Med.*, 1986; 105: 221-233.
- 2 The best biochemical markers of myocardial infarction. *Scripps News*, 1996; 10: 1-4.
- 3 Meerson, F.Z., Javich, M.P., Isoenzyme pattern and activity of myocardial creatine phosphokinase under heart adaptation to prolonged overload. *Basic Res. Cardiol.*, 1982; 77: 349-358.
- 4 Wu, A., Wang, X-M., Gornet, T.G., et al., Creatine kinase MB isoforms in patients with skeletal muscle injury: ramifications for early detection of acute myocardial infarction. *Clin. Chem.*, 1992; 38: 2396-2400.
- 5 Bhayana, V., Cohoe, S., Leung F.Y., et al., Diagnostic evaluation of creatine kinase-2 mass and creatine kinase-3 and -2 isoform ratios in early diagnosis of acute myocardial infarction. *Clin. Chem.*, 1993; 39: 488-495.
- 6 Mair, J., Morandell, D., Genser, N., et al., Equivalent early sensitivities of myoglobin, Creatine Kinase MB mass, Creatine Kinase isoform ratios, and cardiac Troponin I and T for acute myocardial infarction. *Clin. Chem.*, 1995; 41: 1266-1272.
- 7 Bokhari, A.M., Davies, J., Davies J., et al., Biochemical diagnosis of myocardial infarction within the thrombolytic time window. *Int. J. Cardiol.*, 1995; 48: 249-254.
- 8 Mair, J., Wagner, I., Jakob, G., et al., Different time courses of cardiac contractile proteins after acute myocardial infarction. *Clin. Chim. Acta.*, 1994; 231: 47-60.
- 9 Levitt, M.A., Promes, S.B., Bullock, S., et al., Combined cardiac marker approach with adjunct two-dimensional echocardiography to diagnose acute myocardial infarction in the emergency department. *Ann. Em. Med.*, 1996; 27: 1-7.
- 10 Li, D., Jialal, I., Keffer J.H., Greater frequency of increased cardiac Troponin T than increased cardiac Troponin I in patients with chronic renal failure. *Clin. Chem.*, 1996; 42: 114-115.
- 11 Mair, J., Smidt, J., Lechleitner, P., et al., A decision tree for the early diagnosis of acute myocardial infarction in nontraumatic chest pain patient at hospital admission. *Chest*, 1995; 108:1502-1509.
- 12 Yang, Z., Zhang, W., Liu Y., Prognostic efficacy of Troponin T measurement in angina pectoris. *Chin. Med. J. (Engl)*, 1995; 108: 626-630.
- 13 Engvall, E., Enzyme immunoassay ELISA and EMIT. In: Van Vunakis, H. and Langone, J.J. eds., *Methods in Enzymol.*, Academic Press, New York, 1980; 70: 419-439.
- 14 Uotila, M., Ruoslahti, E. and Engvall, E. Two-site sandwich enzyme immunoassay with monoclonal antibodies to human alpha-fetoprotein. *J. Immunol. Methods*, 1981; 42: 11-15.
- 15 Tietz, N.W., ed., *Clinical Guide to Laboratory Tests*, 3rd Edition, W.B. Saunders, Co., Philadelphia, 1995: 180-186.

091503

Oxis International Inc.
323 B Vintage Park Dr. Suite B
Foster City, CA 94404 USA
Phone: 800.547.3686 Fax: 650-212-2568