

Kreatech™ FISH probes **Product Information Sheet**

KBI-10606 BCL2/IGH t(14;18) Fusion













Kreatech Biotechnology B.V. Vlierweg 20 1032 LG Amsterdam The Netherlands www.LeicaBiosystems.com

Published March 2015





Kreatech™ BCL2/IGH t(14;18) Fusion FISH probe

Introduction: Follicular lymphoma is a mature B-Cell lymphoma characterized by the presence of the

t(14;18) translocation that juxtaposes the BCL2 locus on chromosome 18q21 to the immunoglobulin heavy (IGH) locus on chromosome 14q32, resulting in the overexpression of

the antiapoptotic protein BCL2.

Intended use: The BCL2/IGH t(14;18)(q32;q21) specific FISH probe is optimized to detect the reciprocal

translocation t(14;18) in a dual-color, dual-fusion assay on metaphase/interphase spreads,

blood smears and bone marrow cells.

The probe is recommended to be used in combination with one of the Kreatech Pretreatment kits providing necessary reagents to perform FISH on various sample types for optimal

results. (see also www.LeicaBiosystems.com and look for Kits & reagents)

Critical region 1 (red): Critical region 2 (green): The **IGH** (14q32) specific FISH probe is direct-labeled with PlatinumBright[™]550. The **BCL2** (18q21) FISH probe is direct-labeled with PlatinumBright[™]495.

Reagent: Kreatech probes are direct-labeled DNA probes provided in a ready-to-use format.

Apply 10 µl of probe to a sample area of approximately 22 x 22 mm.

Please refer to the Instructions for Use for the entire Kreatech FISH protocol.

Kreatech FISH probes are REPEAT-FREE™ and therefore do not contain Cot-1 DNA. Hybridization efficiency is increased and background, due to unspecific binding, is

highly reduced.

Interpretation: The BCL2/IGH t(14;18) Fusion FISH probe is designed as a dual-fusion probe to detect both

rearranged chromosomes der(18) and der(14) by two co-localized red/green or yellow fusion signals (F). Single color red (R) and green (G) signals will identify the normal chromosomes

14 and 18 respectively.

Signal patterns other than those described above may indicate variant translocations, deletions on der(18) or der(14) or other complex rearrangements. Investigators are advised to analyze metaphase cells for the interpretation of atypical signal patterns.

	Normal Signal Pattern	t(14;18)
Expected Signals	2R2G	2F1R1G

References: Poetsch M et al, 1996, J Clin Oncol, 14; 963- 969

Warning and precautions: In case of emergencies check SDS sheets for medical advice. SDS sheets may be obtained by either contacting Leica Technical Support or visiting www.LeicaBiosystems.com. DNA probes contain formamide which is a teratogen; do not inhale or allow skin contact. Wear gloves and a lab coat when handling DNA probes. All materials should be disposed of according to your institution's guidelines for hospital waste disposal.

Reagent Storage and

Handling:

Store at 2-8 °C. Reagents should not be used after the expiration date on the vial label.

TECHNICAL SUPPORT

Technical support is available at www.LeicaBiosystems.com or +31 20 6919181

or via e-mail: kreatech-support@leicabiosystems.com.

CUSTOMER SERVICE

Kreatech probes may be ordered through Leica Customer Service +31 20 6919181 or order

via e-mail: purchase.orders@leica-microsystems.com.